

Human Thyroid Stimulating Hormone Receptor Antidoby ELISA Kit | TSHRAb elisa kit

Human Thyroid Stimulating Hormone Receptor Antidoby ELISA Kit

CATALOG #
AAA97976

REACTIVITY
Human

PRODUCT SYNONYMS

Thyroid Stimulating Hormone Receptor Antidoby; Human Thyroid Stimulating Hormone Receptor Antidoby ELISA Kit; TSHRAb; TSHRAb elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of TSHRAb. No significant cross-reactivity or interference between TSHRAb and analogues was observed.
Assay Type	Sandwich ELISA, Double Antigen
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	1.563-100mIU/ml
Sensitivity	0.938mIU/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level TSHRAb were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level TSHRAb were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES

Typical Data & Standard Curve

Results of a typical standard run of a TSHRAB ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A/not applicable)

X	mIU/ml	0	1.562	3.125	6.25	12.5	25	50	100
Y	OD450	0.047	0.134	0.197	0.344	0.676	1.078	1.817	2.570

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TSHRAb elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. TSHR was pre-coated onto 96-well plates. And the Biotin- labeled TSHR was used as detection antigen. The standards, test samples and Biotin- labeled TSHR were added to the wells subsequently, and wash with wash buffer. HRP- Streptavidin Conjugate was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the TSHRAb amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of TSHRAb can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.