

Rat Tyrosine Kinase, Endothelial ELISA Kit | TEK elisa kit

Rat Tyrosine Kinase, Endothelial ELISA Kit

CATALOG #
AAA98448

REACTIVITY
Rat


PRODUCT SYNONYMS

Tyrosine Kinase; Endothelial; Rat Tyrosine Kinase; Endothelial ELISA Kit; Tek/CD202b/TEK/angiopoietin-1 receptor/CD202B/CD202b antigen/hTIE2/p140 TEK/soluble TIE2 variant 1/soluble TIE2 variant 2/TEK tyrosine kinase; endothelial/TIE-2/TIE2CD202b/Tunica interna endothelial cell kinase/Tyrosine-protein kinase receptor TEK/Tyrosine-protein kinase receptor TIE-2/venous malformations; multiple cutaneous and mucosal; TEK elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.313-20ng/ml
Sensitivity	0.188ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Typical Data & Standard Curve

Results of a typical standard run of a Rat TEK ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A=not applicable)

X	ng/ml	0	0.313	0.625	1.25	2.5	5	10	20
Y	OD450	0.038	0.118	0.214	0.352	0.588	1.074	1.161	2.116

Standard Curve (Sample)

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TEK elisa kit

Principle of the Assay||This kit was based on sandwich enzyme-linked immune-sorbent assay technology. anti-Rat TEK antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Rat TEK antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Rat TEK amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Rat TEK can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

