

Human Cyclin Dependent Kinase 8 ELISA Kit | CDK8 elisa kit

Human Cyclin Dependent Kinase 8 ELISA Kit

CATALOG #
AAA98465

REACTIVITY
Human


PRODUCT SYNONYMS

Cyclin Dependent Kinase 8; Human Cyclin Dependent Kinase 8 ELISA Kit; CDK8/Protein kinase K35/CDK8 protein kinase/Cell division protein kinase 8/cyclin-dependent kinase 8/K35/Mediator complex subunit CDK8/Mediator of RNA polymerase II transcription subunit CDK8/Protein kinase K35; CDK8 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Sequence Length	462
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.313-20ng/ml
Sensitivity	0.188ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

Typical Data & Standard Curve

Results of a typical standard run of a Human CDK8 ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A-not applicable)

X	ng/ml	0	0.313	0.625	1.25	2.5	5	10	20
Y	OD450	0.145	0.186	0.260	0.294	0.555	0.929	1.475	1.981

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CDK8 elisa kit

Principle of the Assay||This kit was based on sandwich enzyme-linked immune-sorbent assay technology. anti-Human CDK8 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Human CDK8 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Human CDK8 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Human CDK8 can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	351734410
NCBI GeneID	100316862
NCBI Accession #	NP_001162054.2
NCBI GenBank Nucleotide #	NM_001168582.2
NCBI Official Full Name	cyclin dependent kinase 8
NCBI Official Symbol	cdk8
NCBI Protein Information	cyclin dependent kinase 8

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/cdk8-human-elisa-kits/cyclin-dependent-kinase-8/98465>