

Rat Complement Component 4a ELISA Kit | C4a elisa kit

Rat Complement Component 4a ELISA Kit

CATALOG #
AAA98765

REACTIVITY
Rat


PRODUCT SYNONYMS

Complement Component 4a; Rat Complement Component 4a ELISA Kit; C4a/C4A/C4AD/C4S/CO4/CPAMD2/RG/Complement Component 4a; C4a elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.313-20ng/ml
Sensitivity	0.188ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Typical Data & Standard Curve

Results of a typical standard run of a Rat C4a ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A=not applicable)

X	ng/ml	0	0.313	0.625	1.25	2.5	5	10	20
Y	OD450	0.025	0.097	0.190	0.265	0.474	0.903	1.454	2.035

Standard Curve (Sample)

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for C4a elisa kit

Principle of the Assay||This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-Rat C4a antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Rat C4a antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Rat C4a amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Rat C4a can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.