

Mouse Hyaluronan Binding Protein 1 ELISA Kit | HABP1 elisa kit

Mouse Hyaluronan Binding Protein 1 ELISA Kit

CATALOG #
AAA99148

REACTIVITY
Mouse


PRODUCT SYNONYMS

Hyaluronan Binding Protein 1; Mouse Hyaluronan Binding Protein 1 ELISA Kit; HABP1/C1QBP (Hyaluronan Binding Protein 1)/gC1qBP/gC1qR/SF2p32/p32/Glycoprotein gC1qBP/HABP1p33/GC1q-R protein/C1q globular domain-binding protein/gC1Q-R/Hyaluronan-binding protein 1/Mitochondrial matrix protein p32/splicing factor SF2-associated protein; HABP1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.156-10ng/ml
Sensitivity	0.094ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Typical Data & Standard Curve
Results of a typical standard run of a Mouse HABP1 ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A=not applicable)

X	ng/ml	0	0.156	0.313	0.625	1.25	2.5	5	10
Y	OD450	0.025	0.147	0.308	0.457	0.866	1.122	1.818	2.720

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HABP1 elisa kit

Principle of the Assay||This kit was based on sandwich enzyme-linked immune-sorbent assay technology. anti-Mouse HABP1 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Mouse HABP1 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Mouse HABP1 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Mouse HABP1 can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

