

Rat Deoxypridinoline ELISA Kit | DPD elisa kit

Rat Deoxypridinoline ELISA Kit

CATALOG #
AAA99990

REACTIVITY
Rat


PRODUCT SYNONYMS

Deoxypridinoline; Rat Deoxypridinoline ELISA Kit; DPD; DPD elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Assay Type	Competitive ELISA, Coated with Antigen
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	3.125-200ng/ml
Sensitivity	1.875ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Typical Data & Standard Curve

Results of a typical standard run of a Rat DPD ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A=not applicable)

X	ng/ml	0	3.125	6.25	12.5	25	50	100	200
Y	OD450	2.425	1.980	1.653	1.265	0.907	0.596	0.461	0.344

Standard Curve (Sample)

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for DPD elisa kit

Principle of the Assay||This ELISA kit uses Competitive-ELISA as the method. The microtiter plate provided in this kit has been pre-coated with Rat DPD. During the reaction,Rat DPD in the sample or standard competes with a fixed amount of Rat DPD on the solid phase supporter for sites on the Biotinylated Detection Antibody specific to Rat DPD. Excess conjugate and unbound sample or standard are washed from the plate, and HRP-Streptavidin (SABC) is added to each microplate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. The concentration of Rat DPD in the samples is then determined by comparing the O.D. of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.